



[Oxygen Concentrator Recall: JIANGSU JUMAO X-CARE MEDICAL EQUIPMENT CO LTD Removes JMC5A Ni/TruAire-5 Oxygen Concentrator due to the devices spontaneously catching fire | FDA - U.S. Food and Drug Administration](#)

Product Names: TRUAIRE-5 O2 CONCENTRATOR; Model: O2C5L; Serial Numbers: All units within serial number range: JA2311000001-JA2312000740; What to Do

[www.fda.gov](http://www.fda.gov)

For more information on recent events related to concentrators, please visit:

<https://cmscompliancegroup.com/blog/cmscg-client-alert-recent-ny-ct-nursing-home-fires/>